

**3M™ Tegaderm™ + Pad**  
Film Dressing with Non-Adherent Pad

**3M™ Tegaderm™ + Pad**  
Pansement film avec compresse non adhérente

**3M™ Tegaderm™ + Pad**  
Transparentverband mit nicht haftender Wundauflage

**Tegaderm™ + Pad**  
Medicazione trasparente con tampone assorbente

**3M™ Tegaderm™ + pad**  
Apósito transparente con almohadilla antiadherente

**3M™ Tegaderm™ + Pad**  
Filmverband met niet-klevende wondkussen

**3M™ Tegaderm™ + Pad**  
Filmförband med icke-kläbende sårduyna

**3M™ Tegaderm™ + Pad**  
Filmforbindning med ikke-klebende pude

**English**

**French**

**Deutsch**

**Italiano**

**Español**

**Nederlands**

**Svenska**

**Dansk**

**Description:**  
(The 3M™ Tegaderm™ + Pad Film Dressing with Non-Adherent Pad is a waterproof, bacterial and viral\* barrier dressing. The dressing consists of a non-adherent, absorbent pad bonded to a larger thin film backing with a adhesive without natural rubber latex.)  
\*Laboratory testing has proven that Tegaderm provides a viral barrier from viruses 27 nm or larger while the dressing remains intact without leakage.

**Intended Use:**  
The 3M™ Tegaderm™ + Pad Film Dressing with Non-Adherent Pad is intended for covering wounds such as cuts, burns, abrasions, and post-surgical incisions.

**Intended User:**  
For use by medical professionals or under the guidance of medical professionals.

**Contraindications:**  
None known.

**Warnings:**  
Do not use the dressing as a replacement for sutures and other primary wound closure methods.

**Precutions:**  
1. Hemostasis of the catheter site or wound should be achieved before applying the dressing.  
2. Do not stretch the dressing during application as tension can cause skin damage.  
3. Make sure the skin is dry and free of soap residue and lotion to prevent skin irritation and ensure good adhesion.  
4. The dressing may be used on an infected site, only when under the care of a health care professional.

**Directions for Use**  
Do not reuse. Reuse may lead to infection or other illness/injury.  
If the sterile packaging is damaged or unintentionally opened, discard the product, do not use.

**Dressing selection:**  
Choose a dressing size large enough to provide a margin that adheres to dry healthy skin around the catheter or wound site.

**Site preparation:**  
1. Prepare the site according to institution protocol.  
2. Clipping of hair at the site may improve adhesion. Shaving is not recommended.  
3. Allow all prep liquids to completely dry before applying the dressing to prevent skin irritation and ensure good adhesion.

**Application:**  
1. Open package and remove sterile dressing.  
2. Peel the paper liner from the paper-framed dressing, exposing the adhesive surface.  
3. Position the framed window over the wound site or catheter insertion site and apply dressing.  
4. Press the dressing into place.  
5. Remove the paper frame from the dressing while smoothing down the dressing edges. Seal securely around catheter or wound site. Firmly smooth adhesive border to the skin.

**Site Care:**  
1. The site should be observed for signs of infection or other complications. If infection is suspected, remove the dressing, inspect the site directly, and determine appropriate medical intervention.  
2. Infection may be signaled by fever, pain, redness, swelling, or an unusual odor or discharge.  
2. Change the dressing according to institution protocol, or when the barrier properties have been compromised.

**Removal:**  
Gently grasp an edge and slowly peel the dressing from the skin in the direction of hair growth. Avoid skin trauma by peeling the dressing back, rather pulling it up from the skin.

**Shelf Life/Disposal:**  
For shelf life, refer to the expiration date printed on each box. Do not use if expired or if expiration date is missing or illegible. Dispose according to local country legislation.

**How Supplied:**  
Supplied in boxes of individually packaged sterile dressings. Sterility of product is guaranteed unless individual pouch is unintentionally opened or damaged.

**Customer Information:**  
(Please report a serious incident occurring in relation to the device to 3M and the local competent authority (EU) or local regulatory authority. For further information, please contact your local 3M representative or contact us at 3M.com and select your country.)

**MD CE 0297**  
Made in Germany  
■ 3M Deutschland GmbH  
Health Care Business  
Cäcilienstr. 1  
41453 Neuss, Germany  
HCGregulatory 3M.com  
3M and Tegaderm are trademarks of 3M.  
©2023, 3M. Unauthorized use prohibited. All rights reserved.  
Date of issue: 2022-08  
DV-0163-0400-3\_31000472050623

